

labtemp® 勒普拓仪器

:	3
.....	3
.....	4
.....	8
.....	33
.....	35
.....	39
.....	41
.....	50

:	12
---	----

labtemp® 勒普拓仪器




ATEX Directive (2014/34/EU)





70°C.

-30°C.





50 2/

0,1


:

---


HC



1) ( 2)




3) ( 4)




3)

( HC )

---





HC



HC


OBC




( )

( )


- ▶ :
- ▶ 100°C.
- ▶ « HC»!
- ▶ :
- ▶ :
- ▶ ( )
- ▶ :
- ▶ 30
- ▶ :
- ▶ +5 °C +32 °C;
- ▶ :
- ▶ 80%.




- ▶ !
- ▶ :
- ▶ 20
- ▶ 40-60
- ▶ 20-90
- ▶ 90



- ▶ !
- ▶ !
- ▶ !
- ▶ !
- ▶ !










A,

B




B

A,




BC






AM ( )


E / SM



AM ( )


- ▶ 20
- ▶ 40-60
- ▶ 20-90
- ▶ 90







A,

B





A,

B





iCooler






30

+5 °C +32 °C; 80%.




LC1




5°C,  
-20 °C  
( 1:1)

CC

PT100





AM( )





AM( )


**Attention:**

Please refer to the Immersion Heating Circulators section for the E / SM back configuration and more optional




- ▶ 20
- ▶ 40-60
- ▶ 20-90
- ▶ 90









A,




Attention:







Installation:

- ▶ ( )
- ▶ 30
- ▶ +5 °C +32 °C;
- ▶ 80%.




- ▶ 20
- ▶ 40-60
- ▶ 20-100
- ▶ 100





### iM Cooler




▶ ( )

▶ ( )


▶ 30

▶ 80% : +5 °C +32 °C;


▶ ( )



E/SM :




AM :



Number	Name	Function description
A		▶
B	/	▶ 5
C		▶ ( ).
D		▶ ( ).
E	( )	▶ ( 0,1 °C; )
F		▶
G	/	▶ 3
H		▶ ( ).
I	( )	▶ ( 0,1 °C; )
J		▶ SM ( E ).
K	/	▶ ( E, SM ).

8888 → 7.02 → 00dA → H 150

( SM AM):



25°C

		Factory setting
(HIT)	-----	It depends on the model
(LOT)	-----	It depends on the model
(OTH)	-----	It depends on the model
(OTL)	-----	It depends on the model
(BEEP)	0 -----	-
	1 -----	✓
(MODE)	1 -----	✓
	... -----	-
	3 -----	-
(UNIT)	0 -----	-
	1 -----	✓
	... -----	-
(Liquid)	1 -----	-
	... -----	-
	8 -----	✓
(LOCK)	0 -----	✓
	... -----	-
	2 -----	-
(EXT)	0 -----	✓
	1 -----	-
(AUTO)	0 -----	✓
	1 -----	-
(RCAL)	0 -----	✓
	... -----	-
	4 -----	-
(COP)	First point calibration -----	-
	... -----	-
	Third point calibration -----	-
(PARA)	PA12 Kp -----	0.9
	PA13 Ti -----	250
	PA14 Td -----	80

« / »  
 « / », (▲/▼)  
 « / »  
 « / »

(HIT):

HIT

(LOT):

LOT

(OTH):

OTH

«HIT»+10° C.

(OTL):

OTL

«LOT»-10° C.

(BEEP):

BEEP

BEEP = 0:  
 BEEP = 1:

Operation MODE (MODE):

Mode

Mode 1:

Mode 2:

Mode 3:

(UNIT):

UNIT

unit = 0:  
 unit = 1:  
 unit = 2:  
 unit = 3:

(LIQ):

LIQ

		25°C		[ ]	[°C]	
1	TF-M10.80.10	-10...80		90	115	①
2	TF-M30.80.10	-30...80		90	115	②
3	TF-M20.235.20	-20...235	20	245	>200	
4	TF-M40.220.10	-40...220	10	230	>128	
5	TF-M60.200.10	-60...200	10	210	>128	
6	TF-P20.275.50	20...200	50	210	>318	
7	③	5...95		105	-	
8	④					

TF-M40.220.10



① : =1:3

② Glycol: Water = 1:1

③

(LOCK):

LOCK

0:  
 1:  
 2:

(EXT):

EXT

0:  
 1:

(AUTO):

AUTO

0,  
 1,

(AUTO) "PLCK"






00 CIRCULATOR STOP	
01 CIRCULATOR START	
02 PUMP STOP	
03 PUMP START	
00 SENSOR ALARM	
01 SENSOR ALARM	
02 PRESSURE ALARM	
03 COMMUNICATION ALARM	
04 LEVEL ALARM	
05 TEMP ALARM	
06 TEMP ALARM	
07 TEMP ALARM	
08 TEMP ALARM	

Communication	RS485	
Type	A	B
AB lead foot	1	2
Communication distance	Theoretical distance ≤1200m Recommend ≤300m	
Argument	Standard ModBus-RTU	

**RS485**



Baud rate : 19200  
Verify: None  
Data bits: 8  
Stop bits: 1

1. Modbus RTU 16- CRC ( )  
2. 100  
3. : 1 ( 0 100 255)

01	03	00 C8	00 01	CRC
→ ( )				
01	03	02	07 D0	CRC
: 06 →				
01	06	00 C8	07 D0	CRC
( )				
01	06	00 C8	07 D0	CRC

RS485 Modbus RTU 3.1-

( ) ( )			
0x03	0		0,1,
0x03	1		10 ; 0,01, 100 ; 0,001, 1000
0x03/0x06	2		100
0x03/0x06	3	0:	
0x03/0x06	4	1:	
0x03/0x06	5		100 ; 10 320 x 100.
0x03/0x06	6		100 ; -99*100 20*100.
0x03	7	0: 1: 2: 3: 4: 5: 6: 7: 8:	
0x03	8		0.
0x03/0x06	9		50 5000.
0x03	10	1.	0,1, 10 ;
0x03	11	2.	
0x03	12	3.	100 ; 0,01,
0x03	13	4.	0,001, 1000
0x03/0x06	14	0: 1:	
0x03	15		
0x03/0x06	16		
0x03	17		10
0x03/0x06	18		
0x03/0x06	19	1: 2: 3:	
0x03/0x06	20		
0x03/0x06	21	1:	BLDC.
0x03/0x06	22		50 5000.

( ) ( )			
0x03/0x06	23		30 3000.
0x03/0x06	24		10
0x03/0x06	25		0,5*10 5,5*10.
0x03/0x06	26		: 60*10-90*10.
0x03/0x06	27	0: 1:	
0x03/0x06	28		
0x03/0x06	29		
0x03/0x06	30		
0x03/0x06	31		
0x03	32		100
0x03/0x06	33	0: 1:	
0x03/0x06	34		
0x03/0x06	35		1:0.1. 2:0.01. 3:0.001.
0x03/0x06	36		
0x03/0x06	37	0: 1:	
0x03/0x06	38		
0x03/0x06	39	0: 1:	
0x03/0x06	40		10
0x03/0x06	41		
0x03/0x06	42	PID-T.	
0x03/0x06	43	PID-T.	
0x03/0x06	44	PID-Ki	
0x03/0x06	45	PID-Ti.	
0x03/0x06	46	PID-Td.	
0x03/0x06	47		
0x03/0x06	48		
0x03/0x06	49		
0x03/0x06	50		
0x03/0x06	51		
0x03/0x06	52		
0x03/0x06	53		
0x03/0x06	54		

( ) ( )			
0x03/0x06	55		
0x03/0x06	56	ACTDIFF.	100
0x03/0x06	57	4	0:
0x03/0x06	58	4	1:
0x03/0x06	59	4	
0x03/0x06	60	4	0:
0x03/0x06	61		1:
0x03/0x06	62		
0x03/0x06	63		1:
0x03/0x06	64		1:
0x03/0x06	65	NTC	0:PT100. 1:NTC.
0x03/0x06	66		0:
0x03/0x06	67	HWTIME.	1:
0x03	68	MAINTYPE.	
0x03/0x06	69	DEMO.	0:
0x03/0x06	70	PID.	1:
0x03/0x06	71		
0x03	72	4	
0x03	73	4	
0x03/0x06	74		
0x03/0x06	75		
0x03/0x06	76		
0x03/0x06	77		
0x03/0x06	78		
0x03/0x06	79	1.	
0x03/0x06	80	2.	
0x03/0x06	81	3.	
0x03/0x06	82	4.	
0x03/0x06	83	5.	
0x03/0x06	84	6.	
0x03/0x06	85	7.	
0x03/0x06	86	8.	
0x03/0x06	87	9.	
0x03/0x06	88	10.	
0x03/0x06	89	11.	

( ) ( )			
0x03/0x06	90	12.	
0x03/0x06	91	13.	
0x03/0x06	92	14.	
0x03/0x06	93	15.	
0x03/0x06	94	16.	
0x03/0x06	95	17.	
0x03/0x06	96	18.	
0x03/0x06	97	19.	
0x03/0x06	98	20.	
0x03/0x06	99	21.	
0x03	100		
0x03	101		
0x03/0x06	102	485.	
0x03/0x06	103	485.	:0~247.
0x03	104		1:2400 2:4800 3:9600 4:14400 5:19200 6:31250 7:38400 8:57600 9:115200 10:153600 11:230400 12:256000
0x03	105		1: 2: 3:
0x03	106		
0x03	107		
0x03/0x06	108		
0x03/0x06	109		
0x03/0x06	110	CH1.	
0x03/0x06	111	CH2.	
0x03/0x06	112	CH3.	
0x03/0x06	113	CH4.	
0x03/0x06	114	CH5.	
0x03/0x06	115	CH6.	
0x03/0x06	116	CH7.	
0x03/0x06	117	CH8.	
0x03/0x06	118	CH9.	
0x03/0x06	119	CH10.	
0x03/0x06	120	CH11.	
0x03/0x06	121	CH12.	
0x03/0x06	122	CH13.	

( ) ( )			
0x03/0x06	123	CH14.	
0x03/0x06	124	CH15.	
0x03/0x06	125	CH16.	
0x03/0x06	126		0: 1:
0x03/0x06	127		
0x03/0x06	128	1	
0x03/0x06	129	2	
0x03/0x06	130	3	
0x03/0x06	131	4	
0x03/0x06	132	5	
0x03/0x06	133	6	
0x03/0x06	134	7	
0x03/0x06	135	8	
0x03/0x06	136	9	0:Stop. 1:Operation.
0x03/0x06	137	10	
0x03/0x06	138	11	
0x03/0x06	139	12	
0x03/0x06	140	13	
0x03/0x06	141	14	
0x03/0x06	142	15	
0x03/0x06	143	16	
0x03/0x06	144	CH1.	
0x03/0x06	145	CH2.	
0x03/0x06	146	CH3.	
0x03/0x06	147	CH4.	
0x03/0x06	148	CH5.	
0x03/0x06	149	CH6.	
0x03/0x06	150	CH7.	2:2-
0x03/0x06	151	CH8.	
0x03/0x06	152	CH9.	
0x03/0x06	153	CH10.	
0x03/0x06	154	CH11.	
0x03/0x06	155	CH12.	
0x03/0x06	156	CH13.	
0x03/0x06	157	CH14.	
0x03/0x06	158	CH15.	
0x03/0x06	159	CH16.	
0x03	160	CH1	
0x03	161	CH2	
0x03	162	CH3	

( ) ( )			
0x03	163	CH4	
0x03	164	CH5	
0x03	165	CH6	
0x03	166	CH7	
0x03	167	CH8	
0x03	168	CH9	
0x03	169	CH10	
0x03	170	CH11	
0x03	171	CH12	
0x03	172	CH13	
0x03	173	CH14	
0x03	174	CH15	
0x03	175	CH16	
	( )		
0x03	200		0.1, 10 0.01, 100 0.001, 1000
0x03/0x06	230		
0x03/0x06	234		10 320 x 100. 100 ;
0x03/0x06	235		-99*100 20*100. 100 ;
0x03/0x06	236		
0x03/0x06	237		10
0x03/0x06	260		
0x03/0x06	261		0: 1:



labtemp<sup>®</sup> 勒普拓仪器

---

# labtemp<sup>®</sup> 勒普拓仪器

公司名称：深圳勒普拓仪器技术有限公司

公司地址：深圳市龙华区大浪街道新石社区华宁路 42 号 309-310

官方网址：[www.labtemp.cn](http://www.labtemp.cn); [www.labtemp.com.cn](http://www.labtemp.com.cn)

服务电话：4008-266-224

